


PILE DYNAMICS, INC.

GAGE NO. 159D

23-Aug-87



Gage Calibration Factor = 115.0 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -4.8

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.5	.73
3	156.9	1.41
4	233.3	2.10
5	307.6	2.74
6	387.2	3.43
7	463.0	4.08
8	538.7	4.73
9	617.5	5.41
10	693.8	6.07
11	768.7	6.71