


PILE DYNAMICS, INC.

GAGE NO. 160D

23-Aug-87



Gage Calibration Factor = 138.1 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS
 Offset = -5.6

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	80.3	.61
3	158.5	1.20
4	234.1	1.76
5	309.8	2.31
6	384.9	2.86
7	465.1	3.42
8	538.5	3.94
9	618.7	4.52
10	693.1	5.05
11	769.2	5.59