


Pile Dynamics Inc Transducer 162D



PDA Cal Factor = 283.9 ME/V
 Shunt (60.4KΩ) = 5.0 V

General Factor = 985.8 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
131	.46
178	.63
380	1.35
428	1.51
525	1.85
692	2.44
741	2.59
907	3.19
967	3.39
1110	3.93

Calibrated by: J. D. Smith
 Calibrated on: 06-Oct-99