


# PILE DYNAMICS, INC.

GAGE 162W

06-Aug-92



PDA Cal Factor = 176.8 ME/V  
 Ref. Shunt (60.4K) = 5.0 V

General Gage Factor= 614.0 ME/mV/V

| Strain<br>ME | Gage<br>Volts |
|--------------|---------------|
| 0.           | .00           |
| 86.          | .51           |
| 151.         | .90           |
| 224.         | 1.33          |
| 303.         | 1.79          |
| 393.         | 2.30          |
| 468.         | 2.70          |
| 549.         | 3.16          |
| 640.         | 3.66          |
| 706.         | 4.02          |
| 803.         | 4.55          |