


# PILE DYNAMICS, INC.

GAGE NO. 163D

27-Aug-87



Gage Calibration Factor = 109.5 ME/V  
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS  
 Offset = -1.7

| No. | Applied Strain<br>ME | Gage Voltage<br>V |
|-----|----------------------|-------------------|
| 1   | .0                   | .00               |
| 2   | 80.2                 | .74               |
| 3   | 156.1                | 1.45              |
| 4   | 233.1                | 2.15              |
| 5   | 309.6                | 2.85              |
| 6   | 386.7                | 3.55              |
| 7   | 463.8                | 4.24              |
| 8   | 540.6                | 4.95              |
| 9   | 617.6                | 5.64              |
| 10  | 694.7                | 6.37              |
| 11  | 771.2                | 7.03              |