


PILE DYNAMICS, INC. GAGE 165W

07-May-93



PDA Cal Factor = 110.3 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 383.1 ME/mV/V

Strain ME	Gage Volts
0.	.00
50.	.47
106.	.97
164.	1.51
222.	2.03
273.	2.50
326.	2.97
381.	3.46
434.	3.94
491.	4.46
564.	5.03