


PILE DYNAMICS, INC. GAGE 166W

07-May-93



PDA Cal Factor = 173.9 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 603.7 ME/mV/V

Strain ME	Gage Volts
0.	.00
87.	.50
164.	.95
246.	1.43
341.	1.97
434.	2.50
518.	2.98
598.	3.44
689.	3.96
775.	4.46
857.	4.94