


PILE DYNAMICS, INC. GAGE 167W

30-JUN-93



PDA Cal Factor = 121.3 ME/V
Ref. Shunt (60.4K) = 5.0 V

General Gage Factor = 421.3 ME/mV/V

Strain ME	Gage Volts
0.	.00
88.	.73
168.	1.38
261.	2.14
341.	2.79
431.	3.54
509.	4.17
608.	5.00
679.	5.58
761.	6.27
851.	7.02