


PILE DYNAMICS, INC.

GAGE NO. 194D

24-Sep-67



Gage Calibration Factor = 104.5 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS
 Offset = -4.4

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.2	.79
3	156.2	1.53
4	232.9	2.30
5	310.0	3.04
6	389.9	3.82
7	469.5	4.49
8	540.3	5.23
9	618.8	5.94
10	694.5	6.68
11	770.5	7.39