


PILE DYNAMICS, INC.

GAGE NO. 197D

01-Oct-87



Gage Calibration Factor = 103.8 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -4.0

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	80.6	.81
3	156.0	1.54
4	233.7	2.31
5	310.4	3.05
6	387.6	3.80
7	465.6	4.54
8	544.5	5.28
9	619.8	6.00
10	696.9	6.74
11	770.8	7.45