


PILE DYNAMICS, INC.

GAGE NO. 198D

01-Oct-87



Gage Calibration Factor = 126.9 ME/V
 Gage Ref. Shunt (60.4K0) = 5.0 VOLTS
 Offset = -4.7

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.7	.66
3	157.2	1.28
4	234.4	1.91
5	311.1	2.51
6	389.0	3.11
7	463.5	3.70
8	540.9	4.30
9	617.4	4.90
10	696.4	5.52
11	772.9	6.10