


PILE DYNAMICS, INC.

GAGE NO. 235 D

21-Jun-89



PDA Gage Factor = 101.6 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 General Gage Factor = 352.7 ME/mV/V
 Statistical Offset = -3.5

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.3	.79
3	160.2	1.62
4	233.5	2.35
5	309.0	3.10
6	384.6	3.85
7	462.0	4.60
8	538.5	5.35
9	613.1	6.07
10	691.2	6.82
11	769.3	7.58