


PILE DYNAMICS, INC.

GAGE NO. 239D

13-NOV-87



Gage Calibration Factor = 126.6 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -1.9

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.5	.63
3	158.7	1.26
4	234.9	1.88
5	312.6	2.50
6	386.9	3.08
7	463.9	3.69
8	542.5	4.31
9	618.3	4.90
10	693.5	5.49
11	769.2	6.07