


PILE DYNAMICS, INC.

GAGE NO. 241D

13-Nov-87



Gage Calibration Factor = 127.8 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -4.3

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	81.6	.66
3	156.5	1.26
4	233.1	1.88
5	313.1	2.50
6	385.8	3.07
7	462.1	3.67
8	538.1	4.26
9	615.2	4.85
10	694.2	5.46
11	768.3	6.01