


PILE DYNAMICS, INC.

GAGE NO. 242D

13-Nov-87



Gage Calibration Factor = 115.6 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -3.6

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	78.5	.71
3	158.7	1.41
4	234.9	2.08
5	312.6	2.75
6	386.9	3.39
7	463.9	4.05
8	542.5	4.73
9	618.3	5.38
10	693.5	6.02
11	769.2	6.66