


PILE DYNAMICS, INC.

GAGE NO. 243D

13-Nov-87



Gage Calibration Factor = 121.6 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -3.0

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	81.6	.69
3	156.5	1.31
4	233.1	1.96
5	313.1	2.62
6	385.8	3.21
7	462.1	3.84
8	538.1	4.46
9	615.2	5.08
10	694.2	5.72
11	768.3	6.32