


PILE DYNAMICS, INC.

GAGE NO. 245D

13-Nov-87



Gage Calibration Factor = 123.3 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS
 Offset = -4.5

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	77.9	.65
3	152.8	1.27
4	230.3	1.92
5	308.5	2.56
6	384.5	3.18
7	462.1	3.81
8	540.1	4.42
9	614.4	5.01
10	690.6	5.63
11	767.4	6.24