


PILE DYNAMICS, INC.

3/6/86 CAL

GAGE NO. 562 C



INIT

3/6/86 CAL
GAGE NO. 562C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	77.034	0.734
3	156.269	1.470
4	233.340	2.187
5	307.575	2.857
6	377.201	3.466
7	485.224	4.377
8	531.362	4.835
9	606.082	5.486
10	675.970	6.081
11	755.933	6.802

GAGE CALIBRATION FACTOR= 111.723 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -6.5801 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED