


PILE DYNAMICS, INC.

3/20/86 CAL

GAGE NO. 566 C



INIT -----

3/20/86 CAL
GAGE NO. 566C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	80.154	0.699
3	155.722	1.378
4	230.582	2.035
5	303.205	2.680
6	376.946	3.329
7	449.159	3.942
8	526.312	4.628
9	603.334	5.299
10	677.318	5.947
11	755.235	6.616

GAGE CALIBRATION FACTOR= 114.118 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -1.0549 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED