


PILE DYNAMICS, INC.

3/20/86 CAL

GAGE NO. 567 C



INIT -----

3/20/86 CAL
GAGE NO. 567C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	80.005	0.735
3	156.057	1.429
4	230.246	2.092
5	303.615	2.733
6	377.058	3.379
7	447.668	3.992
8	525.995	4.687
9	603.241	5.363
10	676.796	6.015
11	754.191	6.692

GAGE CALIBRATION FACTOR= 113.074 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -3.7894 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED