


PILE DYNAMICS, INC.

3/20/86 CAL

GAGE NO. 568 C



INIT -----

3/20/86 CAL
GAGE NO. 568C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	78.439	0.694
3	155.051	1.409
4	230.321	2.107
5	296.382	2.705
6	380.152	3.463
7	439.727	3.992
8	524.355	4.761
9	603.427	5.476
10	677.579	6.151
11	752.029	6.827

GAGE CALIBRATION FACTOR= 110.132 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -0.1572 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED