


PILE DYNAMICS, INC.

4/21/86 CAL

GAGE NO. 603 C



INIT

4/21/86 CAL
GAGE NO. 603C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	76.295	0.749
3	152.888	1.513
4	229.704	2.267
5	308.268	3.036
6	384.935	3.783
7	461.082	4.525
8	537.935	5.270
9	612.520	6.004
10	687.031	6.732
11	752.878	7.366

GAGE CALIBRATION FACTOR= 102.309 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -0.8639 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED