


PILE DYNAMICS, INC.

4/21/86 CAL

GAGE NO. 604 C



INIT

4/21/86 CAL
GAGE NO. 6040

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	78.117	0.714
3	155.268	1.441
4	231.638	2.155
5	310.294	2.861
6	386.441	3.530
7	462.067	4.190
8	538.418	4.845
9	612.706	5.486
10	686.993	6.123
11	747.468	6.671

GAGE CALIBRATION FACTOR= 112.608 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -6.2036 CORRELATION COEFFICIENT = 0.9998

VALUES ARE AVERAGED