


PILE DYNAMICS, INC.

GAGE NO. 606 C

26-Mar-91



PDA Gage Factor = 102.2 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS
 General Gage Factor = 355.0 ME/mV/V
 Statistical Offset = -4.0

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.4	.78
3	156.0	1.57
4	233.9	2.37
5	306.6	3.08
6	384.0	3.81
7	461.1	4.57
8	539.8	5.33
9	616.7	6.08
10	691.2	6.79
11	768.3	7.51