


PILE DYNAMICS, INC.

GAGE NO. 608 C

27-Mar-91



PDA Gage Factor = 111.8 ME/V
 Gage Ref. Shunt (60.4KΩ) = 5.0 VOLTS

General Gage Factor = 388.2 ME/mv/V
 Statistical Offset = -3.6

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	79.3	.72
3	164.1	1.49
4	239.2	2.21
5	315.1	2.89
6	384.1	3.52
7	462.0	4.21
8	538.9	4.88
9	618.3	5.57
10	689.2	6.19
11	770.9	6.88