


PILE DYNAMICS, INC.

GAGE NO. 609 C

18-Apr-90



PDA Gage Factor		= 101.2 ME/V
Gage Ref. Shunt (60.4KΩ) =		5.0 VOLTS
General Gage Factor		= 351.5 ME/mV/V
Statistical Offset		= -2.4
No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	83.1	.84
3	169.6	1.69
4	234.0	2.35
5	312.7	3.13
6	386.7	3.86
7	473.4	4.73
8	549.2	5.47
9	619.9	6.15
10	692.4	6.85
11	768.8	7.59