


PILE DYNAMICS, INC.

4/24/86 CAL

GAGE NO. 611 C



INIT

4/24/86 CAL
GAGE NO. 611C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	81.646	0.767
3	157.897	1.489
4	229.942	2.179
5	304.294	2.879
6	380.321	3.593
7	456.237	4.304
8	525.677	4.947
9	599.583	5.635
10	674.940	6.338
11	749.665	7.035

GAGE CALIBRATION FACTOR= 107.047 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -1.0635 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED