


PILE DYNAMICS, INC.

GAGE 612C

17-Jan-94



PDA Cal Factor = 121.4 ME/V
 Ref. Shunt (50.4K) = 5.0 V
 General Gage Factor = 421.5 ME/mV/V

Strain ME	Gage Volts
0.	.00
79.	.67
175.	1.48
259.	2.18
338.	2.83
441.	3.68
498.	4.15
589.	4.89
683.	5.65
771.	6.37
854.	7.04