


PILE DYNAMICS, INC.

5/30/86 CAL

GAGE NO. 613 C



INIT

5/30/86 CAL
GAGE NO. 613C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	79.351	0.782
3	156.617	1.544
4	230.569	2.269
5	306.681	3.003
6	380.372	3.702
7	448.329	4.342
8	532.931	5.149
9	605.580	5.841
10	675.659	6.512
11	763.761	7.350

GAGE CALIBRATION FACTOR= 104.169 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -3.3914 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED