


PILE DYNAMICS, INC.

6/3/86 CAL

GAGE NO. 616 C



INIT -----

6/3/86 CAL
GAGE NO. 616C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	80.859	0.735
3	156.822	1.436
4	232.357	2.120
5	307.128	2.807
6	382.662	3.476
7	456.335	4.122
8	530.101	4.768
9	602.005	5.373
10	672.736	6.007
11	750.989	6.676

GAGE CALIBRATION FACTOR= 112.581 ME/V

GAGE REF. SHUNT (60.4KD)= 5.000 VOLTS

OFFSET = -4.5037 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED