


PILE DYNAMICS, INC.

6/3/86 CAL

GAGE NO. 617 C



INIT -----

6/3/86 CAL
GAGE NO. 617C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	79.389	0.752
3	152.931	1.466
4	231.165	2.198
5	311.336	2.933
6	381.229	3.584
7	454.808	4.255
8	528.462	4.918
9	603.755	5.617
10	674.467	6.250
11	751.696	6.954

GAGE CALIBRATION FACTOR= 108.340 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -4.2421 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED