


PILE DYNAMICS, INC.

6/3/86 CAL

GAGE NO. 618 C



INIT

6/3/86 CAL
GAGE NO. 618C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	79.910	0.744
3	152.689	1.451
4	230.271	2.183
5	312.360	2.927
6	380.670	3.566
7	454.250	4.238
8	527.234	4.906
9	604.016	5.589
10	675.678	6.230
11	752.757	6.931

GAGE CALIBRATION FACTOR= 108.739 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -3.9948 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED