


PILE DYNAMICS, INC.

GAGE NO. 620 C

4-19-91



PDA Gage Factor = 103.4 ME/V
 Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 359.1 ME/mV/V
 Statistical Offset = -7.9

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	85.1	.89
3	163.9	1.68
4	239.2	2.42
5	315.7	3.17
6	387.8	3.87
7	467.2	4.64
8	545.5	5.36
9	621.5	6.09
10	697.8	6.80
11	772.8	7.50