


PILE DYNAMICS, INC.

6/4/86 CAL

GAGE NO. 621 C



INIT -----

6/4/86 CAL
GAGE NO. 621C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	77.859	0.657
3	154.006	1.292
4	230.525	1.915
5	310.393	2.536
6	396.435	3.202
7	469.384	3.771
8	537.719	4.315
9	613.197	4.909
10	680.789	5.424
11	759.131	6.041

GAGE CALIBRATION FACTOR= 126.304 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -6.6665 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED