


PILE DYNAMICS, INC.

6/6/86 CAL

GAGE NO. 628 C



INIT

6/6/86 CAL
GAGE NO. 628C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	81.596	0.780
3	157.475	1.505
4	228.568	2.138
5	309.103	2.917
6	379.209	3.543
7	454.064	4.216
8	525.456	4.864
9	598.468	5.526
10	678.145	6.245
11	753.168	6.918

GAGE CALIBRATION FACTOR= 109.154 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -4.7811 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED