


PILE DYNAMICS, INC.

6/20/86 CAL

GAGE NO. 639 C



INET

6/20/86 CAL
GAGE NO. 639C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	82.199	0.786
3	157.961	1.483
4	231.695	2.185
5	306.880	2.870
6	381.470	3.541
7	456.525	4.205
8	528.547	4.866
9	602.746	5.526
10	676.294	6.177
11	753.321	6.869

GAGE CALIBRATION FACTOR= 110.221 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -5.2992 CORRELATION COEFFICIENT = 0.9999

VALUES ARE AVERAGED