


PILE DYNAMICS, INC.

6/20/86 CAL

GAGE NO. 640 C



INIT

6/20/86 CAL
GAGE NO. 640C

NO.	APPLD. STRAIN ME	GAGE VOLTAGE V
1	0.000	0.000
2	78.404	0.811
3	154.612	1.596
4	232.793	2.389
5	305.131	3.105
6	378.214	3.821
7	454.273	4.563
8	532.454	5.324
9	601.927	6.013
10	676.796	6.755
11	752.558	7.490

GAGE CALIBRATION FACTOR= 100.835 ME/V

GAGE REF. SHUNT (60.4K0)= 5.000 VOLTS

OFFSET = -4.8014 CORRELATION COEFFICIENT = 1.0000

VALUES ARE AVERAGED