


PILE DYNAMICS, INC.

GAGE NO. 641 C

20-Jul-89



PDA Gage Factor = 102.8 ME/V
Gage Ref. Shunt (60.4K) = 5.0 VOLTS

General Gage Factor = 357.1 ME/mV/V
Statistical Offset = -1.8

No.	Applied Strain ME	Gage Voltage V
1	.0	.00
2	74.6	.74
3	159.8	1.57
4	226.0	2.22
5	310.9	3.05
6	390.4	3.83
7	462.9	4.53
8	536.8	5.25
9	612.4	5.98
10	691.3	6.73
11	763.3	7.42