


PILE DYNAMICS, INC.

GAGE 460H

17-JUL-93



PDA Cal Factor = 104.0 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 361.1 ME/mV/V

Strain ME	Gage Volts
0.	.00
89.	.87
148.	1.44
218.	2.11
289.	2.79
368.	3.54
435.	4.19
508.	4.88
582.	5.59
680.	6.55
731.	7.05