


PILE DYNAMICS, INC.

GAGE 461H

17-JUL-93



PDA Cal Factor = 105.7 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 366.9 ME/mV/V

Strain ME	Gage Volts
0.	.00
77.	.74
148.	1.40
235.	2.23
309.	2.93
376.	3.56
453.	4.29
513.	4.85
592.	5.59
672.	6.36
743.	7.04