


PILE DYNAMICS, INC.

GAGE

467H

27-Jul-93



PDA Cal Factor = 107.9 ME/V
 Ref. Shunt (50.4KΩ) = 5.0 V
 General Gage Factor = 374.7 ME/mV/V

Strain ME	Gage Volts
0.	.00
77.	.72
150.	1.40
234.	2.18
303.	2.82
380.	3.52
469.	4.34
537.	4.97
606.	5.62
686.	6.36
762.	7.08