


PILE DYNAMICS, INC.

GAGE 469H

27-Jul-93



PDA Cal Factor = 105.0 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 364.6 ME/mV/V

Strain ME	Gage Volts
0.	.00
73.	.70
147.	1.41
221.	2.10
295.	2.81
367.	3.50
446.	4.25
513.	4.88
592.	5.64
667.	6.35
733.	6.99