


PILE DYNAMICS, INC.

GAGE 478H

27-JUL-93



PDA Cal Factor = 105.1 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 364.8 ME/mV/V

Strain ME	Gage Volts
0.	.00
78.	.75
148.	1.42
223.	2.13
301.	2.86
372.	3.54
448.	4.26
516.	4.91
589.	5.60
665.	6.33
734.	7.00