


PILE DYNAMICS, INC.

GAGE 479H

27-Jul-93



PDA Cal Factor = 104.6 ME/V
 Ref. Shunt (50.4KQ) = 5.0 V
 General Gage Factor = 363.1 ME/mV/V

Strain ME	Gage Volts
0.	.00
71.	.69
142.	1.36
213.	2.05
292.	2.79
366.	3.50
436.	4.17
509.	4.87
581.	5.56
661.	6.32
724.	6.94