


PILE DYNAMICS, INC.

GAGE

482H

27-JUL-93



PDA Cal Factor = 103.5 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 359.3 ME/mV/V

Strain ME	Gage Volts
0.	.00
75.	.73
143.	1.38
222.	2.15
290.	2.80
362.	3.50
430.	4.16
510.	4.92
581.	5.62
648.	6.26
724.	7.00