


PILE DYNAMICS, INC.

GAGE 483H

27-JUL-93



PDA Cal Factor = 109.0 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 378.4 ME/mV/V

Strain ME	Gage Volts
0.	.00
79.	.73
151.	1.38
239.	2.19
307.	2.81
380.	3.48
464.	4.25
537.	4.92
612.	5.61
689.	6.32
765.	7.03