


PILE DYNAMICS, INC.

GAGE 486H

28-Jul-93



PDA Cal Factor = 104.3 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 362.2 ME/mV/V

Strain ME	Gage Volts
0.	.00
71.	.68
158.	1.52
219.	2.10
298.	2.86
365.	3.49
442.	4.23
519.	4.97
582.	5.57
661.	6.34
730.	7.01