


PILE DYNAMICS, INC.

GAGE 489H

28-JUL-93

PDA Cal Factor = 103.9 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 360.8 ME/mV/V



Strain ME	Gage Volts
0.	.00
74.	.71
144.	1.38
234.	2.26
300.	2.86
381.	3.66
435.	4.18
517.	4.97
588.	5.66
651.	6.27
731.	7.04