


PILE DYNAMICS, INC.

GAGE 490H

28-JUL-93



PDA Cal Factor = 103.5 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 359.4 ME/mV/V

Strain ME	Gage Volts
0.	.00
74.	.71
148.	1.43
215.	2.07
292.	2.82
364.	3.51
436.	4.21
505.	4.87
578.	5.58
646.	6.24
721.	6.97