


PILE DYNAMICS, INC.

GAGE

491H

28-Jul-93



PDA Cal Factor = 104.6 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 363.2 ME/mV/V

Strain ME	Gage Volts
0.	.00
93.	.88
149.	1.43
223.	2.13
295.	2.82
378.	3.60
445.	4.25
521.	4.97
586.	5.60
659.	6.30
731.	7.00